

राष्ट्रीय डोप परीक्षण प्रयोगशाला
भारत सरकार

जवाहर लाल नेहरू स्टेडियम खेल परिसर
पूर्वी द्वार सं. 10, लोधी रोड
नई दिल्ली – 110003

दिनांक : 02.03.2026

File No. M-12/10/2024-Admin

CORRIGENDUM

Bid No. GEM/2026/B/71787, Dated 03.02.2026 for procurement of 01 Unit of LCMS System - Below Mentioned Point Shall Be Read as Follows And Deemed To Be replaced from the Original Bid Document:-

S. No.	Floated Specification	Amendments shall be read as
1	At S. No. 15. <u>Detector</u> Latest technology Pulse Counting CEM/PEM/PMT detector to ensure wide dynamic linear range. Dynamic range of at least 6 orders of magnitude.	"Latest technology Pulse Counting CEM/PEM/PMT/EMT detector to ensure wide dynamic linear range. Dynamic range of at least 6 orders of magnitude."

This issues with the approval of the Competent Authority.



(अंशु सक्सेना)

उप-निदेशक (प्रशासन)